



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

A Uno Calibration Laboratory S.A.
CondominioTERRUM local 25, Río Segundo, Alajuela, Costa Rica

*(Hereinafter called the Organization) and hereby declares that Organization is accredited
in accordance with the recognized International Standard:*

ISO/IEC 17025:2017

This accreditation demonstrates technical competence for a defined scope and the
operation of a laboratory quality management system
(as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

***Chemical, Dimensional, Electrical, Mass (Weighing Devices, Individual
Weight), Mechanical, Thermodynamic, Time & Frequency Calibrations
(As detailed in the supplement)***

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen
President

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

Initial Accreditation Date: April 9, 2008 *Issue Date:* August 17, 2024 *Expiration Date:* December 31, 2026

Accreditation No.: 59381 *Certificate No.:* L24-634

*The validity of this certificate is maintained through ongoing assessments based on a
continuous accreditation cycle. The validity of this certificate should be
confirmed through the PJLA website: www.pilabs.com*



Certificate of Accreditation: Supplement

A1 Calibration Laboratory S.A.

TERRUM 25 Condominium, Rio Segunda, Alajuela, Costa Rica
Contact Name: Mr. Felix Hernandez Phone: 506-2440-4010

Accreditation is granted to the facility to perform the following calibration:

Chemical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
pH Meters ^{FO}	Up to 14 pH	0.03 pH	Buffers Solutions	Method I-30

Dimensional

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Caliper ^{FO}	Up to 600 mm	(0.007 L + 6) μ m	Gage Blocks	Method I-03
Micrometer ^{FO} (inside)	Up to 305 mm	(0.007 L + 6) μ m		Method I-09
Micrometer ^{FO} (outside)	Up to 400 mm	(0.007 L + 6) μ m		Method I-18
Depth Micrometer ^{FO}	Up to 150 mm	(0.007 L + 6) μ m		Method I-39
Dial Indicator ^{FO}	Up to 50 mm	(0.007 L + 6) μ m	Gage Blocks	Method I-02
Steel Rule ^F	Up to 1 000 mm	0.06 mm	Gage Block Portable microscope OMAX M51B	Method I-01
Pin Gages ^F	Up to 25.4 mm	0.002 mm	Laser Micrometer	Method I-04
Measuring Tape ^F	Up to 12 000 mm	0.8 mm	Standard Steel Rule	Method I-01
Protractor ^{FO}	0° to 90°	0.1 °	Angle Gage Blocks	Method I-12 Method I-35

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Temperature Calibration, Indication and Control Equipment used with Thermocouple Type T ^{FO}	0 °C to 400 °C	0.7 °C	Process Calibrator 753 8.5 DMM 8104	Method I-14.3
Temperature Calibration, Indication and Control Equipment used with Thermocouple Type R ^{FO}	50 °C to 1 700 °C	1.2 °C		
Temperature Calibration, Indication and Control Equipment used with Thermocouple Type J ^{FO}	-210 °C to 750 °C	0.5 °C		



Certificate of Accreditation: Supplement

A1 Calibration Laboratory S.A.

TERRUM 25 Condominium, Rio Segunda, Alajuela, Costa Rica
Contact Name: Mr. Felix Hernandez Phone: 506-2440-4010

Accreditation is granted to the facility to perform the following calibration:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Temperature Calibration, Indication and Control Equipment used with Thermocouple Type K ^{FO}	-140 °C to 1 340 °C	0.4 °C	Process Calibrator 753 8.5 DMM 8104	Method I-14.3
Temperature Calibration Indication & Control Equipment used RTD Type Pt 385, 100 Ω ^{FO}	-200 °C to 800 °C	0.23 °C	Process Calibrator 753	Method I-14.5
Temperature Calibration Indication & Control Equipment used RTD Type Pt 3926, 100 Ω ^{FO}	-200 °C to 630 °C	0.12 °C		
Temperature Calibration Indication & Control Equipment used RTD Type Pt 3916, 100 Ω ^{FO}	-200 °C to 360 °C	0.25 °C		
Temperature Calibration Indication & Control Equipment used RTD Type Pt 385, 200 Ω ^{FO}	-200 °C to 630 °C	0.16 °C		
Temperature Calibration Indication & Control Equipment used RTD Type Pt 385, 500 Ω ^{FO}	-200 °C to 630 °C	0.12 °C		
Temperature Calibration Indication & Control Equipment used RTD Type Pt 385, 1000 Ω ^{FO}	-200 °C to 630 °C	0.23 °C		
Temperature Calibration Indication & Control Equipment used RTD Type Pt Ni 672, 120 Ω ^{FO}	-200 °C to 260 °C	0.14 °C		
Temperature Calibration Indication & Control Equipment used RTD Type Cu 427, 10 Ω ^{FO}	-100 °C to 260 °C	0.3 °C		
Temperature Calibration Indication & Control Equipment used RTD Type Pt 3926, 100 Ω ^{FO}	-200 °C to 630 °C	0.5 °C		
Temperature Calibration Indication & Control Equipment used RTD Type Pt 385, 100 Ω ^{FO}	-200 °C to 800 °C	0.8 °C		



Certificate of Accreditation: Supplement

A1 Calibration Laboratory S.A.

TERRUM 25 Condominium, Rio Segunda, Alajuela, Costa Rica
Contact Name: Mr. Felix Hernandez Phone: 506-2440-4010

Accreditation is granted to the facility to perform the following calibration:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Temperature Calibration Indication & Control Equipment used RTD Type Pt 385, 200 Ω ^{FO}	-200 °C to 630 °C	0.8 °C	Process Calibrator 753	Method I-14.5
Temperature Calibration Indication & Control Equipment used RTD Type Pt 385, 500 Ω ^{FO}	-200 °C to 630 °C	0.8 °C	8.5 DMM 8104	Method I-11 Method I-19 Method I-38
Equipment to Output DC Voltage ^{FO}	0 mV to 100 mV	5.8 μ V/V + 0.9 μ V		
	100 mV to 1 V	4.6 μ V/V + 2.5 μ V		
	1 V to 10 V	4.8 μ V/V + 4.5 μ V		
	10 V to 100 V	8.5 μ V/V + 72 μ V		
	100 V to 1 000 V	25 μ V/V + 250 μ V		
Equipment to Output AC Voltage at the listed Frequencies ^{FO}				
10 Hz to 40 Hz	0 mV to 100 mV	86 μ V/V + 99 μ V		
40 Hz to 200 Hz	0 mV to 100 mV	86 μ V/V + 45 μ V		
200 Hz to 2 kHz	0 mV to 100 mV	86 μ V/V + 38 μ V		
2 kHz to 20 kHz	0 mV to 100 mV	86 μ V/V + 53 μ V		
20 kHz to 100 kHz	0 mV to 100 mV	86 μ V/V + 160 μ V		
Equipment to Output AC Voltage at the listed Frequencies ^{FO}				
10 Hz to 40 Hz	100 mV to 1 V	92 μ V/V + 50 μ V		
40 Hz to 200 Hz	100 mV to 1 V	92 μ V/V + 31 μ V		
200 Hz to 2 kHz	100 mV to 1 V	170 μ V/V + 31 μ V		
2 kHz to 20 kHz	100 mV to 1 V	350 μ V/V + 31 μ V		
20 kHz to 100 kHz	100 mV to 1 V	930 μ V/V + 31 μ V		
100 kHz to 1 MHz	100 mV to 1 V	3.5 mV/V + 120 μ V		
Equipment to Output AC Voltage at the listed Frequencies ^{FO}				
10 Hz to 40 Hz	1 V to 10 V	92 μ V/V + 50 μ V		
40 Hz to 200 Hz	1 V to 10 V	92 μ V/V + 31 μ V		
200 Hz to 2 kHz	1 V to 10 V	170 μ V/V + 31 μ V		
2 kHz to 20 kHz	1 V to 10 V	350 μ V/V + 31 μ V		
20 kHz to 100 kHz	1 V to 10 V	930 μ V/V + 31 μ V		
100 kHz to 200 kHz	1 V to 10 V	3.5 mV/V + 120 μ V		



Certificate of Accreditation: Supplement

A1 Calibration Laboratory S.A.

TERRUM 25 Condominium, Rio Segunda, Alajuela, Costa Rica
Contact Name: Mr. Felix Hernandez Phone: 506-2440-4010

Accreditation is granted to the facility to perform the following calibration:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Equipment to Output AC Voltage at the listed Frequencies ^{FO}			8.5 DMM 8104	Method I-11 Method I-19 Method I-38
10 Hz to 40 Hz	10 V to 100 V	450 μ V/V + 2.7 mV		
40 Hz to 200 Hz	10 V to 100 V	450 μ V/V + 2.7 mV		
200 Hz to 2 kHz	10 V to 100 V	450 μ V/V + 2.7 mV		
2 kHz to 20 kHz	10 V to 100 V	560 μ V/V + 2.7 mV		
20 kHz to 50 kHz	10 V to 100 V	1.5 mV/V + 2.7 mV		
Equipment to Output AC Voltage at the listed Frequencies ^{FO}				
10 Hz to 40 Hz	100 V to 1 000 V	450 μ V/V + 2.7 mV		
40 Hz to 200 Hz	100 V to 1 000 V	450 μ V/V + 2.7 mV		
200 Hz to 2 kHz	100 V to 1 000 V	560 μ V/V + 2.7 mV		
2 kHz to 20 kHz	100 V to 1 000 V	1.5 mV/V + 2.7 mV		
20 kHz to 50 kHz	100 V to 1 000 V	4.7 mV/V + 2.7 mV		
Equipment to Output DC Current ^{FO}				
1 nA to 10 nA	4 μ A/A + 0.2 nA			
10 nA to 100 nA	8 μ A/A + 0.8 nA			
100 nA to 1 μ A	12 μ A/A + 1.2 nA			
1 μ A to 10 μ A	24 μ A/A + 1.2 nA			
10 μ A to 100 μ A	24 μ A/A + 1 nA			
100 μ A to 1 mA	24 μ A/A + 7.1 nA			
1 mA to 10 mA	24 μ A/A + 69 nA			
10 mA to 100 mA	41 μ A/A + 680 nA			
100 mA to 1 A	130 μ A/A + 13 μ A			
1 A to 10 A	130 μ A/A + 26 μ A			
10 A to 30 A	130 μ A/A + 80 μ A			
Equipment to Output AC Current at the listed Frequencies ^{FO}				
10 Hz to 40 Hz	0 μ A to 100 μ A	0.47 % of Reading + 35 nA		
40 Hz to 1 kHz	0 μ A to 100 μ A	0.18 % of Reading + 35 nA		
1 kHz to 10 kHz	0 μ A to 100 μ A	0.07 % of Reading + 35 nA		
Equipment to Output AC Current at the listed Frequencies ^{FO}				
10 Hz to 40 Hz	100 μ A to 1 mA	0.47 % of Reading + 35 nA		
40 Hz to 1 kHz	100 μ A to 1 mA	0.18 % of Reading + 35 nA		
1 kHz to 10 kHz	100 μ A to 1 mA	0.07 % of Reading + 35 nA		
Equipment to Output AC Current at the listed Frequencies ^{FO}				
10 Hz to 40 Hz	1 mA to 10 mA	0.18 % of Reading + 240 nA		
40 Hz to 1 kHz	1 mA to 10 mA	0.07 % of Reading + 240 nA		
1 kHz to 10 kHz	1 mA to 10 mA	0.035 % of Reading + 240 nA		



Certificate of Accreditation: Supplement

A1 Calibration Laboratory S.A.

TERRUM 25 Condominium, Rio Segunda, Alajuela, Costa Rica
Contact Name: Mr. Felix Hernandez Phone: 506-2440-4010

Accreditation is granted to the facility to perform the following calibration:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED		
Equipment to Output AC Current at the listed Frequencies ^{FO}			8.5 DMM 8104	Method I-11 Method I-19 Method I-38		
10 Hz to 40 Hz	10 mA to 100 mA	0.18 % of Reading + 38 μ A				
40 Hz to 1 kHz	10 mA to 100 mA	0.07 % of Reading + 43 μ A				
1 kHz to 10 kHz	10 mA to 100 mA	0.035 % of Reading + 24 μ A				
Equipment to Output AC Current at the listed Frequencies ^{FO}						
10 Hz to 40 Hz	100 mA to 1 A	0.47 % of Reading + 240 μ A				
40 Hz to 1 kHz	100 mA to 1 A	0.19 % of Reading + 240 μ A				
1 kHz to 10 kHz	100 mA to 1 A	0.12 % of Reading + 240 μ A				
Equipment to Output AC Current at the listed Frequencies ^{FO}						
10 Hz to 40 Hz	1 A to 10 A	0.13 % of Reading + 150 μ A				
40 Hz to 1 kHz	1 A to 10 A	0.1 % of Reading + 150 μ A				
Equipment to Output AC Current at the listed Frequencies ^{FO}						
10 Hz to 40 Hz	10 A to 30 A	0.16 % of Reading + 1.5 mA	8.5 DMM 8104	Method I-37 Method I-32		
40 Hz to 1 kHz	10 A to 30 A	0.11 % of Reading + 1.5 mA				
Equipment to Output Resistance ^{FO}	0 Ω to 1 Ω	22 μ Ω / Ω + 80 μ Ω				
	1 Ω to 10 Ω	18 μ Ω / Ω + 130 μ Ω				
	10 Ω to 100 Ω	14 μ Ω / Ω + 1.1 m Ω				
	100 Ω to 1 k Ω	12 μ Ω / Ω + 1 m Ω				
	1 k Ω to 10 k Ω	12 μ Ω / Ω + 7.4 m Ω				
	10 k Ω to 100 k Ω	12 μ Ω / Ω + 320 m Ω				
	100 k Ω to 1 M Ω	18 μ Ω / Ω + 7.1 Ω				
	1 M Ω to 10 M Ω	59 μ Ω / Ω + 46 Ω				
	10 M Ω to 100 M Ω	58 μ Ω / Ω + 5.7 k Ω				
	100 M Ω to 1 G Ω	58 μ Ω / Ω + 0.11 M Ω				
	1 G Ω to 10 G Ω	58 μ Ω / Ω + 130 k Ω				
	10 G Ω to 100 G Ω	58 μ Ω / Ω + 1.1 M Ω				
	100 G Ω to 1 T Ω	58 μ Ω / Ω + 11 M Ω				



Certificate of Accreditation: Supplement

A1 Calibration Laboratory S.A.

TERRUM 25 Condominium, Rio Segunda, Alajuela, Costa Rica
Contact Name: Mr. Felix Hernandez Phone: 506-2440-4010

Accreditation is granted to the facility to perform the following calibration:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Equipment to Output Capacitance ^{FO}	0 nF to 1 nF	2 % of Reading + 0.025 nF	8.5 DMM 8104	Method I-11 Method I-19 Method I-38
	1 nF to 10 nF	1 % of Reading + 0.05 nF		
	10 nF to 100 nF	1 % of Reading + 0.5 nF		
	100 nF to 1 μ F	1 % of Reading + 5 nF		
	1 μ F to 10 μ F	1 % of Reading + 50 nF		
	10 μ F to 100 μ F	1 % of Reading + 0.5 μ F		
	100 μ F to 1 mF	1 % of Reading + 5 μ F		
	1 mF to 10 mF	1 % of Reading + 50 μ F		
	10 mF to 100 mF	1 % of Reading + 0.2 mF		
Hypot Tester ^{FO}	100 V to 9 000 V	280 μ V/V + 15 mV	Vitrek 4700	Method I-45
pH Meters ^{FO}	-2 000 mV to 2 000 mV	0.3 mV	Fluke 753	Method I-11
	15 °C to 40 °C	0.5 °C	Fluke 753/ K Thermocouple	Method I-14

Mass, Force, and Weighing Devices

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Balance ^{FO}	0 g to 100 g	(0.45 + 0.003Wt) mg	Class F1 Weight Set	Method I-13
	101 g to 300 g	(0.62 + 0.003Wt) mg		
	301 g to 1 000 g	(1.2 + 0.003Wt) mg		
	1 001 g to 2 000 g	(10 + 0.003Wt) mg		
	2 001 g to 10 000 g	(150 + 0.003Wt) mg		
	10 001 g to 30 000 g	(320 + 0.003Wt) mg		
	30 001 g to 60 000 g	(36 + 0.003Wt) g		
	60 001 g to 425 kg	(210 + 0.003Wt) g		
Weight (Mass) ^{FO}	0.5 g	0.06 mg	OIML F1, F2, M1, M2	CEM Weight Calibration Procedure Method I-27
	1 g	0.07 mg		
	2 g	0.08 mg		
	5 g	0.1 mg		
	10 g	0.14 mg		
	20 g	0.16 mg		
	50 g	0.24 mg		
	100 g	0.46 mg		



Certificate of Accreditation: Supplement

A1 Calibration Laboratory S.A.

TERRUM 25 Condominium, Rio Segunda, Alajuela, Costa Rica
Contact Name: Mr. Felix Hernandez Phone: 506-2440-4010

Accreditation is granted to the facility to perform the following calibration:

Mass, Force, and Weighing Devices

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Weight (Mass) ^{FO}	200 g	0.92 mg	OIML F2, M1, M2	CEM Weight Calibration Procedure Method I-27
	500 g	2.4 mg		
	1 000 g	4.6 mg		
	2 000 g	9.6 mg		
	5 000 g	24 mg		
	10 000 g	46 mg		
	20 000 g	92 mg		
Liquid Volume Measuring Devices to include Graduated Cylinders, Beakers, Burets, Erlenmeyer, Glass Micro Pipettes, Volumetric Balls, Imhoff Cones, Seraphin Test Measures, Gallon "to contain container or bucket" ^F	(Up to 200) mL	0.003 mL	DIGITAL BALANCE OHAUS PA2202 DIGITAL BALANCE METTLER TOLEDO MS32001LE	Gravimetric Method Method CENAM Technical Guide
	(200 to 6 000) ml	0.46 mL		
	(6 000 to 25 000) mL	2.8 mL		

Mechanical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Pressure Gauge, Differential Pressure Indicators, Sensors and Transducer ^{FO}	0 kPa to 2.49 kPa	0.014 kPa	Ashcroft ATE-2/AM2-1 Transducer	Method I-06
	2.49 kPa to 206 kPa	0.042 kPa		
Vacuum Gauges, Vacuum transducers and Sensors ^{FO}	-75.8 kPa to 0 kPa	0.042 kPa	DPI 150	Method I-07
Pressure Gauge, Sensors and Transducer ^{FO}	206 kPa to 2 068 kPa	0.18 kPa	UPM Module	
	2 068 kPa to 3 447 kPa	0.41 kPa		
Pressure Gauge S ^{FO}	0 kPa to 6 894 kPa	0.40 kPa	Module 700P30	
	6 894 kPa to 20 684 kPa	0.56 kPa		
	20 684 kPa to 34 473 kPa	4.5 kPa		
	34 473 kPa to 68 948 kPa	8.0 kPa	Module DP 104	
Torque Wrench ^{FO}	0.5 Nm to 5.6 Nm	0.75 % of Reading	Transducer 2000-400-2	Method I-16
	3.4 Nm to 45.2 Nm	0.75 % of Reading		
	9.0 Nm to 113 Nm	0.75 % of Reading		
	27.1 Nm to 339 Nm	0.75 % of Reading		
	271.2 Nm to 2 711 Nm	0.75 % of Reading	Transducer 2000-14-02	



Certificate of Accreditation: Supplement

A1 Calibration Laboratory S.A.

TERRUM 25 Condominium, Rio Segunda, Alajuela, Costa Rica
Contact Name: Mr. Felix Hernandez Phone: 506-2440-4010

Accreditation is granted to the facility to perform the following calibration:

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Volume Delivery Instruments (Pipettes) ^{FO}	(20 to 100) μ L	0.8 μ L	MICROBALANCE RADWAG BALANCE & SCALES XA 21.4Y.M.A PLUS ANALITICAL BALANCE OHAUS PA224	Method ISO 8655-6
	(100 to 1 000) μ L	1.8 μ L		
	(1 000 to 5 000) μ L	8.9 μ L		
	(5 000 to 10 000) μ L	18 μ L		

Thermodynamic

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
RTD Thermometer ^{FO}	-25 °C to 420 °C	0.08 °C	PTR 5615 / 1521	Method I-14.4
	250 °C to 420 °C	0.1 °C		
Temperature Bath ^{FO}	-25 °C to 300 °C	0.15 °C	PTR 5613 / 1521	Method I-14.1
Liquid in Glass Thermometer, Digital Thermometer, Bimetallic Thermometer ^{FO}	-25 °C to 300 °C	0.3 °C	PTR 5615 / 1521 Dry Well 650 S Liquid Bath TE-10D	Method I-14.2 Method I-14 Method I-14.3
Oven, Freezer, Furnace, Digital Thermometer with thermocouple. ^{FO}	-25 °C to 650 °C	1.7 °C	Fluke 753 / TC type J and K / Graphtec GL 220 / Dry Well 650 S	Method I-14.1
Digital Infrared Thermometer ^{FO}	-25 °C to 35 °C	1.8 °C	Liquid Bath TE-10D with blackbody target with TC type K	Method I-15
	35 °C to 400 °C	0.3 °C		
Thermo Hygrometer ^{FO}	15 °C to 30 °C	0.3 °C	Standard Thermo Hygrometer 635-1	Method I-42
	20 % RH to 95 % RH	2.5 % RH		

Time & Frequency

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Time and Frequency Simulation ^{FO}	1 μ Hz to 80 MHz	3×10^{-5} Hz/Hz	Function Generator 4086	Method I-46 Method I-40 Method I-21
	Up to 2.4 GHz	60×10^{-6} Hz/Hz		
RPM Digital Measurement Instrument ^{FO}	10 rpm to 90 000 rpm	0.05 % of Reading	Function Generator 4086	Method I-43



Certificate of Accreditation: Supplement

A1 Calibration Laboratory S.A.

TERRUM 25 Condominium, Rio Segunda, Alajuela, Costa Rica
Contact Name: Mr. Felix Hernandez Phone: 506-2440-4010

Accreditation is granted to the facility to perform the following calibration:

1. The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor k (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
2. The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.
3. The presence of a superscript F means that the laboratory performs calibration of the indicated parameter at its fixed location.
4. The presence of a superscript O means that the laboratory performs calibration of the indicated parameter onsite at customer locations.
5. Measurement uncertainties obtained for calibrations performed at customer sites can be expected to be larger than the measurement uncertainties obtained at the laboratories fixed location for similar calibrations. This is due to the effects of transportation of the standards and equipment and upon environmental conditions at the customer site which are typically not controlled as closely as at the laboratories fixed location